



Certificate of Compliance

Certificate: 1934001 (LR65844C)

Master Contract: 176045

Project: 1934001

Date Issued: 2007/07/16

Issued to: Carl Zeiss Jena GmbH
Pruefzentrum
Carl-Zeiss-Promenade 10
Jena, 07745
Germany
Attention: Andreas Holz

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Peter Cheung, P. Eng.

Authorized by: Renzo Pupulin, C.E.T., Product Group Manager

PRODUCTS

CLASS 8721 85 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

CLASS 8721 05 - LABORATORY EQUIPMENT - Electrical

Laser scanning microscope; Model LSM 5 EXCITER; rated 120/230 V; 50/60Hz; 3,6 kVA; Pollution Degree 2; Installation Category II: IP20

APPLICABLE REQUIREMENTS

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1934001 (LR65844C)

Master Contract: 176045

Project: 1934001

Date Issued: 2007/07/16

CAN/CSA-C22.2 No. 61010-1-04 Safety Requirements for Electrical Equipment, Control, and Laboratory Use -
Part I - General Requirements.

UL 61010-1:2004 2nd Edition Safety Requirements for Electrical Equipment for Measurement, Control and
Laboratory Use - Part I - General Requirements.

MARKINGS



Supplement to Certificate of Compliance

Certificate: 1934001

Master Contract: 176045

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1934001	2007/07/16	Laser Scanning Microscope model LSM 5 EXCITER

History

Supplement Notes